



Technical note

Evaluation of the accuracy of musculoskeletal simulation during squats by means of instrumented knee prostheses[☆]

Florian Schellenberg^a, William R. Taylor^{a,*}, Adam Trepczynski^b, Renate List^a, Ines Kutzner^b, Pascal Schütz^a, Georg N. Duda^b, Silvio Lorenzetti^{a,c}

^aInstitute for Biomechanics, ETH Zurich, Leopold-Ruzicka-Weg 4, 8093 Zürich, Switzerland

^bJulius Wolff Institute, Charité – Universitätsmedizin Berlin, Germany

^cSwiss Federal Institute of Sport Magglingen, SFISM, Magglingen, Switzerland

ARTICLE INFO

Article history:

Received 28 November 2017

Revised 8 August 2018

Accepted 16 September 2018

Keywords:

Joint contact force

Deep knee flexion

Musculoskeletal modelling

In vivo validation

Strength training

OpenSim

ABSTRACT

Standard musculoskeletal simulation tools now offer widespread access to internal loading conditions for use in improving rehabilitation concepts or training programmes. However, despite broad reliance on their outcome, the accuracy of such loading estimations, specifically in deep knee flexion, remains generally unknown. The aim of this study was to evaluate the error of tibio-femoral joint contact force (JCF) calculations using musculoskeletal simulation compared to *in vivo* measured JCFs in subjects with instrumented total knee endoprostheses during squat exercises.

Using the early but common “Gait2392_simbody” (OpenSim) scaled musculoskeletal models, tibio-femoral JCFs were calculated in 6 subjects for 5 repetitions of squats. Tibio-femoral JCFs of 0.8–3.2 times bodyweight (BW) were measured. While the musculoskeletal simulations underestimated the measured knee JCFs at low flexion angles, an average error of less than 20% was achieved between approximately 25°–60° knee flexion. With an average error that behaved almost linearly with knee flexion angle, an overestimation of approximately 60% was observed at deep flexion (ca. 80°), with an absolute maximum error of ca. 1.9BW. Our data indicate that loading estimations from early musculoskeletal gait models at both high and low knee joint flexion angles should be interpreted carefully.

© 2018 The Author(s). Published by Elsevier Ltd on behalf of IPPEM.

This is an open access article under the CC BY-NC-ND license.

(<http://creativecommons.org/licenses/by-nc-nd/4.0/>)

1. Introduction

Accurate knowledge of the internal loading conditions in the human musculoskeletal (MS) system, including muscle and joint contact forces (JCFs), can provide a strong evidence-based foundation for improving rehabilitation concepts and customising training programmes, as well as for optimising implant designs. Although direct, non-invasive access to muscle and JCFs is, in most cases, not possible, internal loading conditions have become widely available using MS simulation [1]. The accuracy of such simulations, especially if large ranges of motion (RoMs) are considered, is known to be sensitive to a wide variety of parameters, but often differs in a complex manner from the real *in vivo* situation [2,3].

During strength exercises, extreme postures with large joint angles such as high knee flexion during squats is of interest since the greatest joint loading conditions are thought to occur in these postures [3,4]. While modelling approaches to estimate these internal muscle and JCFs have become widely available in a variety of software modelling packages (e.g., Anybody, OpenSim, LifeModeler™, MSIM, Biomechanics of Bodies, etc.), validation of the output of MS simulations, especially at higher joint flexion [5], remains difficult due to limited *in vivo* data. One risk is the growing reliance on knowledge gained from standard reference models for application to a wide range of subjects and activities, including rehabilitation and training exercises with high knee flexion angles, without knowing the associated validity of data and the corresponding levels of error. In selected cases, *in vivo* force measurements using instrumented total knee endoprostheses are possible [6], allowing access to the JCFs during a variety of activities, thus providing a reference for evaluating musculoskeletal load analyses. In the recent CAMS-Knee project, such telemetric measurements of *in vivo* knee joint loading were combined with

[☆] The authors would also like to acknowledge that this work was included in the Ph.D. thesis of Dr Florian Schellenberg, who agrees with the inclusion of these data in this manuscript.

* Corresponding author.

E-mail address: bt@ethz.ch (W.R. Taylor).




Fig. 1. Experimental set-up. One participant performing a squat within the c-arm of the moving fluoroscope.

detailed analyses of skeletal motion using moving videofluoroscopy in 6 subjects [7]. Using these unique data, the aim of this study was to evaluate the accuracy of JCFs determined using open source MS simulation tools compared to tibio-femoral JCFs measured in 6 subjects each with an instrumented total knee arthroplasty (TKA) during squat exercises.

2. Methods

This kinematics and kinetics data used in this study were based on the CAMS-Knee (www.cams-knee.orthoload.com) datasets. While the measurement protocols are described in the literature [7], a brief description of the data is provided here: six subjects (5 m, 1 f, aged 68 ± 5 years, mass 88 ± 12 kg, height 173 ± 4 cm) were measured while performing five repetitions of a squat exercise without additional weight. Each subject possessed an INNEX knee replacement (Zimmer, Switzerland; type FIXUC), in which the tibial component was instrumented with a 9-channel telemetry transmitter (90–100 Hz) that allowed six-component load measurements of the 3 JCFs and 3 joint moments acting on the tibial component to be recorded within an accuracy of 2% [6]. To assess the motion of the body, 55 skin markers were attached mainly to the lower extremities [8], and an opto-electronic system (Vicon, Oxford Metrics Group, UK) with 22 cameras (MX40 and MX160) captured the kinematics (100 Hz). The ground reaction forces (GRFs) were measured using two force plates (Kistler, Switzerland; 2 kHz), one under each foot. The CAMS-Knee study was approved by the local ethics committees of the Charité (EA4/069/06) and ETH Zürich (EK 2013-N-90) and all subjects provided written informed consent prior to participation.

Each subject performed basic motion tasks to functionally determine the centres (fCoRs) and axes of rotation of the hip, knee




Fig. 2. Musculoskeletal model [OpenSim SimTK 3.3, Stanford, USA; 9] based on the common “Gait2392_simbody” model [10–15] including 14 body segments, 23 degrees of freedom and 92 muscles, shown for an exemplary subject during the squat exercise.

and ankle joints [8]. The squat activity then consisted of each subject standing with stationary feet, approximately shoulder width apart, starting with the eccentric phase (Fig. 1). Knee joint flexion was then performed as far as possible before returning to the standing position. The kinematic and kinetic data were reconstructed in Vicon Nexus (v1.8.5, Oxford Metrics Group, UK) and further processed using in-house software written using Matlab (R2014a, Mathworks, USA) to extract skin marker and joint centre locations for each time frame, as well as joint angles and GRFs.

To estimate tibio-femoral JCFs, MS simulations of each subject's lower limbs were created that included scaling to their segment lengths (determined from the fCoRs), inverse kinematics and inverse dynamics to determine the joint moments, and finally optimization to determine the muscle forces, as described below. The models [OpenSim SimTK 3.3, Stanford, USA; 9] were based on the common “Gait2392_simbody” model [10–15] including 14 body segments, 23 degrees of freedom and 92 muscles [16] (Fig. 2) without residual reduction analysis [17]. Each model's segment lengths, as well as muscles' slack and tendon lengths, were scaled based on the fCoRs of the ankle, knee and hip [9]. To reconstruct the

Table 1

Extreme (max or min) joint contact forces (JCF) of the knee measured in subjects using instrumented total knee arthroplasty (TKA) or calculated by means of musculoskeletal simulation (MS) as well as the Averaged Peak Error [%].

Subject	Instrumented implant		MS simulation		Averaged Peak Error	
	$JCF_{TKA,min}$ [BW]	$JCF_{TKA,max}$ [BW]	$JCF_{MS,min}$ [BW]	$JCF_{MS,max}$ [BW]	Error JCF_{min} [%]	Error JCF_{max} [%]
1	1.16	2.50	0.53	3.95	−54.1	58.2
2	0.86	2.54	0.44	4.20	−48.9	65.5
3	0.75	1.97	0.59	3.32	−21.0	68.6
4	0.94	3.21	0.87	5.11	−7.5	59.0
5	0.88	2.13	0.47	3.42	−47.3	60.7
6	1.34	3.25	0.59	4.52	−55.8	39.2
Ave.	0.99	2.60	0.58	4.09	−39.1	58.5
Std	0.20	0.49	0.14	0.62	18.2	9.4

subjects' kinematics, marker weightings were chosen as follows: hip, knee and ankle with 50; skin markers with 1, and an automated weighting procedure based on soft tissue artefacts, with a total weight of 10 for each segment [18–20].

To calculate muscle moment arms, especially for multi-joint muscles, the OpenSim generalized force method was used, which takes wrapping and via points of muscles into account [21]. A static optimization criteria that minimized the sum of the squared Hill-type muscle [16,22] activation was used [23–25]. Total simulated JCFs (JCF_{MS}) were calculated as the sum of the muscle, ground reaction, segment inertial forces, and masses.

To evaluate the accuracy of the quasi-static optimization approach for determining internal forces at the knee, JCF_{MS} were compared against the measured JCFs from the instrumented TKAs (JCF_{TKA}). The Error JCF was calculated for each trial as follows:

$$\text{Error JCF [in \%]} = \frac{(JCF_{MS} - JCF_{TKA})}{JCF_{TKA}} \cdot 100 \quad (1)$$

Additionally, the results obtained for all 6 subjects were arithmetically averaged inter-individually and presented as a function of the knee flexion angle [°]. Furthermore, the maximal and minimal measured ($JCF_{TKA,max}$, $JCF_{TKA,min}$) and calculated ($JCF_{MS,max}$, $JCF_{MS,min}$) forces within each cycle were extracted and used to calculate the average peak error (Error JCF_{max}) between the simulated and measured JCFs.

3. Results

JCFs on the tibial plateau (JCF_{TKA}) of between 0.8 and 3.2 times bodyweight (BW) were measured *in vivo* across the six patients during the squat activity (Table 1 and Fig. 3). The simulated forces for the same activities were calculated as 0.4–5.1 BW, both under- and over-estimating the measured JCF_{TKA} at different flexion angles. The averaged peak error (Error JCF_{max}) over all subjects and squat cycles predicted by the MS models was 58.5%, while the inter-subject standard deviations of this error were around 9% (Fig. 4). The associated RMS error of the simulated joint centres and marker positions were 10 mm and 13 mm respectively. The average total model residual force (at the pelvis) was 19.6 N.

An almost linear dependence of the Error JCF was observed with knee flexion, resulting in single cycle absolute errors of up to 100% at deep knee flexion angles (Fig. 4). The range of knee flexion where the errors in the JCF remained between $\pm 20\%$ was approximately 25°–60° during both eccentric and concentric movements.

4. Discussion

The usage of reference models for the determination of internal loading conditions has become commonplace, but the accuracy of such gait models, particularly during other more challenging activities that include higher RoMs has, until now, remained




Fig. 3. Joint contact forces (in BW) as measured using the instrumented implant are presented for each subject (thin dashed coloured lines) together with the estimated JCFs from the MS models for each subject (continuous lines; same colour as the measured forces) over the complete range of flexion throughout the squat activity. The mean JCF is shown as a thick line.




Fig. 4. Errors in the joint contact force (error JCF, shown in %) of the simulated forces are shown compared to the telemetrically measured knee joint contact force across the range of knee flexion achieved during the squat activity. Since not all participants reached the same range of knee flexion, data from only the region where 4 or more trials were available is presented. The thin lines represent the average error in each individual subject, shown in subject specific colours consistent with Fig. 3. The black thick line depicts the average of all 6 subjects (averaged JCF error). The grey areas show the average range of joint flexion (dashed lines; 26.5°–60° eccentric; 27.5°–60.5° concentric) that achieved an averaged JCF error of within $\pm 20\%$.

Table 2

Muscle moment arms [cm] around the hip and knee joints for different muscles (m. gluteus maximus, hamstrings, quadriceps, and m. gastrocnemius) at the time points where maximal JCF-errors (under- and over-estimated) occur. Note that some of the lever arms could be considered to be extremely small, thus providing a subjective indication of possible sources of JCF error. Note that the moment arms of all muscles are larger at the time point of JCF underestimation than overestimation.

		Moment Arms at max over-estimation [cm]	Moment Arms at max under-estimation [cm]
Hip	glutMax1	-1.9 ± 0.5	-5.3 ± 0.7
	glutMax2	-2.4 ± 0.6	-5.9 ± 0.7
	glutMax3	-4.1 ± 0.9	-6.6 ± 0.2
	glutMed1	0.0 ± 0.4	-2.7 ± 0.8
	glutMed3	-2.0 ± 0.3	-3.4 ± 0.6
	glutMin1	-0.2 ± 0.3	-1.6 ± 0.7
	glutMin3	-0.7 ± 0.2	-2.2 ± 0.7
	biFemLh	-4.5 ± 0.9	-5.9 ± 0.6
	semimem	-4.2 ± 0.8	-5.4 ± 0.5
	semiten	-5.4 ± 0.9	-6.3 ± 0.8
	rectFem	4.4 ± 0.8	4.7 ± 0.6
	biFemLh	-1.9 ± 0.4	-3.4 ± 0.1
	biFemSh	-2.4 ± 0.4	-3.1 ± 0.2
Knee	semimem	-3.1 ± 0.4	-3.6 ± 0.2
	semiten	-3.6 ± 0.5	-4.4 ± 0.3
	rectFem	3.0 ± 0.3	5.2 ± 0.2
	vasInt	3.0 ± 0.2	4.9 ± 0.2
	vasLat	3.1 ± 0.2	4.7 ± 0.2
	vasMed	3.2 ± 0.2	4.7 ± 0.2
	latGas	-1.4 ± 0.2	-4.2 ± 0.3
	medGas	-1.6 ± 0.1	-4.0 ± 0.3

limited. In this study, a reference open source MS model has been bench-marked against *in vivo* measured forces for six subjects throughout repetitions of the squat activity. Our data suggest a clear relationship between the error in force calculation and the knee flexion angle. However, with an average error range of $\pm 20\%$ (individual errors of up to $+53\%$ or -45%), the results of this study do indicate that the investigated MS model is indeed able to estimate knee JCFs during activities where the knee joint flexion angle remains within approximately 25° – 60° . These data are in agreement with previously observed JCFs during the late stance phase of gait (25% – 100% stance phase) but differ from the first 25% , where an overestimation of the maximal JCF has been reported [2,26]. In our study, the lack of impact, shear forces, and one-legged stance phases during squats, which are all present during gait, are likely to play key roles in explaining the observed underestimation of JCFs, and this is consistent with known problems associated with transferability between different activities [27,28]. As a result, adaptation of the results during squatting to other (weight-bearing) activities might be limited and should be handled with care. At high flexion, the lever arms of some muscles were observed to be particularly low (Table 2), possibly resulting in an unfavourable redistribution of the forces throughout other muscles, and well might contribute the large overestimation of JCFs. A large inter-subject variation was also observed (Table 1, Figs. 3 and 4), signifying that results of a single individual should be interpreted with care.

During strength training, the highest loading conditions are associated with positive adaptation but also possible injury, and often occur in positions of deep knee flexion [4]. However, the loading conditions in these postures are now known to be estimated with least accuracy and should therefore be interpreted with caution. The flexion dependent error also possibly explains the larger average peak errors ($Error_{JCF_{max}}$) found in our study (58.5%) compared to a previous study analysing daily activities such as gait (11%) or stair climbing (26%) [5], where their activities were performed at lower knee flexion angles. However, since a low peak tibio-femoral JCF error of $14 \pm 10\%$ was observed in their study at $\sim 90^\circ$ flexion and peak tibio-femoral contact forces of 2.2 – 2.3 BW were simulated by others [29] during squatting activities, it seems

that the error is likely to be governed by more than just differences in joint flexion angles alone.

While it is entirely possible that a regression relationship ($Error_{JCF} [\%] = 1.3 \cdot x + 57.5$; where x represents the knee flexion angle; $R^2 = 0.99$), could be used to correct predictions when using the Gait2392_simbody model, it should be recognised that systematic improvements to the underlying model should be targeted if possible, rather than adjusting the output forces to reduce errors by means of a correction factor. The clear question posed by the results of our study is: why does the investigated standard MS model have less capacity to predict forces at extreme ranges of knee flexion? Based on the nature of the squat movement, the external knee joint moment increases almost linearly with the flexion angle [4]. This means that the demand on the extensor muscles increases with flexion, suggesting that outside of the $\sim 25^\circ$ – 60° range, the muscle loading is either excessively low or high. It is therefore probable that an improvement to subject specific anatomical parameters [26], dynamic wrapping paths and lever arms of muscles [particularly at the knee and hip; [30,31], as well as the consideration of passive soft- and connective-tissue forces (e.g., ligaments, joint capsule etc.), muscle co-contraction [29], muscle and tendon slack lengths [32], and goal-oriented muscle optimisation [26,33], will all play key roles for achieving improved predictions at high and low flexion angles. In this study, only a subjective assessment of the sources of error could be performed (Table 2), but further investigations to better identify the true origins of the errors in an objective manner should systematically examine the relative importance of each parameter using e.g., Monte Carlo analyses.

One of the strengths of this study was the usage of the complete CAMS-Knee datasets [7] as a basis for the musculoskeletal analyses performed. These datasets combine state-of-the-art measurements of internal joint contact forces at the knee [6,34] with skeletal motion of the knee joint using moving videofluoroscopy [35] and whole-body kinematics using motion capture [36], and will be made available for public release at <https://cams-knee.orthoload.com>. As a result, the data driving the analysed models is the current gold standard. However, there are a number of limitations to this study to be considered. It is clear that our results are based only on a small population of elderly

subjects, using implants where the cruciate ligaments have been sacrificed [3,34]. In this respect, the lever arm of the patella tendon in particular, which is known to play a critical role for the prediction of tibio-femoral JCFs [26], might well differ from the standard, and relatively old version of the Opensim MS models used in our analyses due to the implanted patella button. However, it is important to emphasise that more up-to-date MS models are now available. The next stages of this work will therefore establish whether more detailed models are able to reliably predict JCFs.

In this study, a knee joint flexion dependent error of < 20% was observed between approximately 25°–60°, suggesting that widely available MS models can allow access to loading predictions at mid-range flexion angles, but indicates that loading estimations should be interpreted carefully. An improved understanding of the aetiology of the potentially large errors is clearly required.

Acknowledgements

We would like to acknowledge the financial support of the Robert Mathys Foundation (RMS) for enabling the measurement of the subjects with telemetric knee implants to occur at ETH Zürich, Switzerland. The authors would also like to thank Verena Schwachmeyer, Barbara Postolka, Joern Dymke, Marco Hitz and Michael Angst for their help during the data collection. Additional funding was also provided by the German Federal Ministry of Education and Research (OVERLOAD-PrevOP, 01EC1408A).

Conflicts of interest

There are no conflicts of interest.

Funding

We would like to acknowledge the financial support of the Robert Mathys Foundation (RMS) for enabling the measurement of the subjects with telemetric knee implants (within the original CAMS-Knee study: <http://cams-knee.orthoload.com/>) to occur at ETH Zürich, Switzerland.

Ethical approval

The data used in this study were captured from subjects measured in a previous investigation: the CAMS-Knee project, which was approved by the local ethics committees of the Charité (EA4/069/06) and ETH Zürich (EK 2013-N-90). All subjects provided written informed consent prior to participation.

References

- [1] Schellenberg F, Oberhofer K, Taylor WR, Lorenzetti S, Lorenzetti S. Review of modelling techniques for *in vivo* muscle force estimation in the lower extremities during strength training. *Comput Math Methods Med* 2015;2015.
- [2] Fregly BJ, Besier TF, Lloyd DG, Delp SL, Banks SA, Pandy MG, et al. Grand challenge competition to predict *in vivo* knee loads. *J Orthop Res: Off Publ Orthop Res Soc* 2012;30:503–13.
- [3] Bergmann G, Bender A, Graichen F, Dymke J, Rohlmann A, Trepczynski A, et al. Standardized loads acting in knee implants. *PLoS One* 2014;9:e86035.
- [4] Lorenzetti S, Gülay T, Stoop M, List R, Gerber H, Schellenberg F, et al. Comparison of the angles and corresponding moments in the knee and hip during restricted and unrestricted squats. *J Strength Cond Res* 2012;26:2829–36.
- [5] Trepczynski A, Kutzner I, Kornaropoulos E, Taylor WR, Duda GN, Bergmann G, et al. Patellofemoral joint contact forces during activities with high knee flexion. *J Orthop Res* 2012;30:408–15.
- [6] Heinlein B, Graichen F, Bender A, Rohlmann A, Bergmann G. Design, calibration and pre-clinical testing of an instrumented tibial tray. *J Biomech* 2007;40:S4–S10.
- [7] Taylor WR, Schutz P, Bergmann G, List R, Postolka B, Hitz M, et al. A comprehensive assessment of the musculoskeletal system: the CAMS-Knee data set. *J Biomech* 2017;65:32–9.
- [8] List R, Gülay T, Stoop M, Lorenzetti S. Kinematics of the trunk and the lower extremities during restricted and unrestricted squats. *J Strength Cond Res* 2013;27:1529–38.
- [9] Delp SL, Anderson FC, Arnold AS, Loan P, Habib A, John CT, et al. OpenSim: open-source software to create and analyze dynamic simulations of movement. *IEEE Trans Biomed Eng* 2007;54:1940–50.
- [10] Anderson FC, Pandy MG. Dynamic optimization of human walking. *J Biomech Eng* 2001;123:381–90.
- [11] Anderson FC, Pandy MG. A dynamic optimization solution for vertical jumping in three dimensions. *Comput Methods Biomech Biomed Eng* 1999;2:201–31.
- [12] Delp SL, Loan JP, Hoy MG, Zajac FE, Topp EL, Rosen JM. An interactive graphics-based model of the lower extremity to study orthopaedic surgical procedures. *Biomed Eng IEEE Trans* 1990;37:757–67.
- [13] Hamner SR, Seth A, Delp SL. Muscle contributions to propulsion and support during running. *J Biomech* 2010;43:2709–16.
- [14] Debaere S, Delecluse C, Aerenhouts D, Hagman F, Jonkers I. Control of propulsion and body lift during the first two stances of sprint running: a simulation study. *J Sports Sci* 2015;33:2016–24.
- [15] Navacchia A, Rullkoetter PJ, Schutz P, List RB, Fitzpatrick CK, Shelburne KB. Subject-specific modeling of muscle force and knee contact in total knee arthroplasty. *J Orthop Res: Off Publ Orthop Res Soc* 2016.
- [16] Thelen DG. Adjustment of muscle mechanics model parameters to simulate dynamic contractions in older adults. *J Biomech Eng* 2003;125:70–7.
- [17] Schellenberg F. Muscle forces in the lower extremities during strength training for strength enhancement, prevention and rehabilitation of an anterior cruciate ligament rupture. ETH Zürich; 2016.
- [18] Heller MO, Kratzstein S, Ehrig RM, Wassilew G, Duda GN, Taylor WR. The weighted optimal common shape technique improves identification of the hip joint center of rotation *in vivo*. *J Orthop Res* 2011;29:1470–1475.
- [19] Kratzstein S, Kornaropoulos EI, Ehrig RM, Heller MO, Pöplau BM, Taylor WR. Effective marker placement for functional identification of the centre of rotation at the hip. *Gait Posture* 2012;36:482–6.
- [20] Schellenberg F, Taylor WR, Jonkers I, Lorenzetti S. Robustness of kinematic weighting and scaling concepts for musculoskeletal simulation. *Comput Methods Biomech Biomed Eng* 2017;1–10.
- [21] Sherman MA, Seth A, Delp SL. What is a moment arm? Calculating muscle effectiveness in biomechanical models using generalized coordinates. In: *Proceedings of the ASME 2013 International Design Engineering Technical Conferences and Computers and Information in Engineering Conference*. American Society of Mechanical Engineers; 2013 V07BT10A052–V07BT10A.
- [22] Yamaguchi GT, Zajac FE. A planar model of the knee joint to characterize the knee extensor mechanism. *J Biomech* 1989;22:1–10.
- [23] Rasmussen J, Damsgaard M, Voigt M. Muscle recruitment by the min/max criterion—a comparative numerical study. *J Biomech* 2001;34:409–15.
- [24] Herzog W. Individual muscle force estimations using a non-linear optimal design. *J Neurosci Methods* 1987;21:167–79.
- [25] Happee R. Inverse dynamic optimization including muscular dynamics, a new simulation method applied to goal directed movements. *J Biomech* 1994;27:953–60.
- [26] Gerus P, Sartori M, Besier TF, Fregly BJ, Delp SL, Banks SA, et al. Subject-specific knee joint geometry improves predictions of medial tibiofemoral contact forces. *J Biomech* 2013;46:2778–86.
- [27] Shelburne KB, Pandy MG. A dynamic model of the knee and lower limb for simulating rising movements. *Comput Methods Biomech Biomed Eng* 2002;5:149–59.
- [28] Wu W, Lee PV, Bryant AL, Galea M, Ackland DC. Subject-specific musculoskeletal modeling in the evaluation of shoulder muscle and joint function. *J Biomech* 2016;49:3626–34.
- [29] Mizu-uchi H, Colwell CW, Flores-Hernandez C, Fregly BJ, Matsuda S, D'Lima DD. Patient-specific computer model of dynamic squatting after total knee arthroplasty. *The J Arthroplast* 2015;30:870–4.
- [30] Arnold AS, Salinas S, Asakawa DJ, Delp SL. Accuracy of muscle moment arms estimated from MRI-based musculoskeletal models of the lower extremity. *Comput Aided Surg: Off J Int Soc Comput Aided Surg* 2000;5:108–119.
- [31] Smith SM, Cockburn RA, Hemmerich A, Li RM, Wyss UP. Tibiofemoral joint contact forces and knee kinematics during squatting. *Gait Posture* 2008;27:376–86.
- [32] Modenese L, Ceseracciu E, Reggiani M, Lloyd DG. Estimation of musculotendon parameters for scaled and subject specific musculoskeletal models using an optimization technique. *J Biomech* 2016;49:141–8.
- [33] Yamasaki T, Idehara K, Xin X. Estimation of muscle activity using higher-order derivatives, static optimization, and forward-inverse dynamics. *J Biomech* 2016;49:2015–22.
- [34] Kutzner I, Heinlein B, Graichen F, Bender A, Rohlmann A, Halder A, et al. Loading of the knee joint during activities of daily living measured *in vivo* in five subjects. *J Biomech* 2010;43:2164–73.
- [35] List R, Postolka B, Schutz P, Hitz M, Schwilch P, Gerber H, et al. A moving fluoroscope to capture tibiofemoral kinematics during complete cycles of free level and downhill walking as well as stair descent. *PLoS One* 2017;12:e0185952.
- [36] Taylor WR, Kornaropoulos EI, Duda GN, Kratzstein S, Ehrig RM, Arampatzis A, et al. Repeatability and reproducibility of OSSCA, a functional approach for assessing the kinematics of the lower limb. *Gait Posture* 2010;32:231–6.